

Nurse News

Global EpiPen Recall

Mylan has announced a voluntary recall of EpiPen and EpiPen Jr auto-injectors. The recall is due to a defective part that may result in the device failing to inject the life-saving medication epinephrine.

A list of affected products follows. ***All student EpiPens at school have been checked and parents have been notified.***

If you think you may be impacted by this recall, it is very important that you first contact Stericycle at 877-650-3494. Stericycle's hours of operation are Monday-Friday 8 a.m.-10 p.m. ET, and Saturday and Sunday 8 a.m.-5 p.m. ET.

Product/Dosage	NDC Number	Lot Number	Expiration Date
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	5GN767	April 2017
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	5GN773	April 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	5GM631	April 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	5GM640	May 2017
EpiPen Jr Auto-Injector, 0.15 mg	49502-501-02	6GN215	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM082	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM072	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM081	September 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM088	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM199	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM091	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM198	October 2017
EpiPen Auto-Injector, 0.3 mg	49502-500-02	6GM087	October 2017

References:

<http://www.mylan.com/epipenrecall>

<https://www.fda.gov/newsevents/newsroom/pressannouncements/ucm550170.htm>